

**Visual Acuity Lighting Log**

**Purpose:** To record light meter readings regarding visual acuity lane and chart lighting.  
**When:** Monthly.  
**By Whom:** Visual Acuity Examiner.  
**Instructions:** Lane and chart light readings should be taken on a monthly basis. Start at the 10 foot (or 4m) floor marker in the visual acuity lane. Hold the meter approximately 4 feet from the floor with the light sensor on top of the meter facing the ceiling for all readings. Take a second reading at the 5 foot (or 2m) mark. The third reading should be made directly in front of the chart.

Date of reading (day-month-year):	Lane reading (ft. candles):		Chart reading (ft. candles):	Visual acuity examiner signature:
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			
	10' (4m)			
	5' (2m)			